Department of Health & Social Care

Date: 22/10/2021

Important Customer Notice- Quarantine and Recall

Dear Colleague

Obisk Blue Tree Face Type IIR mask - MKTIIR0214 (tie back) and MKTIIR0199 (ear looped). Batch/LOT numbers:

- **101000345**
- **101000346**
- **101000348**
- **101000349**
- **101000350**

Our records show that may you have received these items since July of this year. Please note, any items with a manufacture date prior to July 2021 are not affected by this notice.

These are labelled Type IIR, however information received by the manufacturer indicates that the masks **may not** meet technical specifications for splash protection. This affects only the batch numbers stated above.

The manufacturer has submitted a report to the MHRA.

MKTIIR0199

MKTIIR0214

Department of Health & Social Care

Indication of where to locate LOT number on inner box/ carton

Action Required

Pending further investigation, please do not use in clinical environments. Please conduct an immediate review of your mask stock, in order to identify if you have any of the affected products.

If you have any of these products, please inform the NHS Product Enquiries Team productqueries@ppeenquiries.com, quoting **Ref: Obisk Blue Tree Face Type IIR mask** and confirm the volume of product in stock by **15/11/21**. We will then give further instructions determined by service environment, please also let us know if replacement stock is required.

Further Actions
The manufacturer has instigated a full investigation, overseen by the regulator. Once we have further information, an update will be provided

Yours faithfully,

Moira Philpott
Chair Decision Making Committee